

2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma)

Package Insert

A rapid test for the qualitative detection of IgG and IgM antibodies to 2019-nCoV in human whole blood, serum or plasma specimens.

For professional *in vitro* diagnostic use only.

【INTENDED USE】

The 2019-nCoV IgG/IgM Rapid Test Cassette is a lateral flow chromatographic immunoassay for the qualitative detection of IgG and IgM antibodies to 2019-nCoV in human whole blood, serum or plasma specimens.

【SUMMARY】

Early January 2020, a novel coronavirus (2019-nCoV) was identified as the infectious agent causing an outbreak of viral pneumonia in Wuhan, China, where the first cases had their symptom onset in December 2019.¹

Coronaviruses are enveloped RNA viruses that are distributed broadly among humans, other mammals, and birds and that cause respiratory, enteric, hepatic, and neurologic diseases.² Six coronavirus species are known to cause human disease.³ Four viruses — 229E, OC43, NL63, and HKU1 — are prevalent and typically cause common cold symptoms in immunocompromised individuals.³ The two other strains — severe acute respiratory syndrome coronavirus (SARS-COV) and Middle East respiratory syndrome coronavirus (MERS-COV) — are zoonotic in origin and have been linked to sometimes fatal illness.⁴ Coronaviruses are zoonotic, meaning they are transmitted between animals and people.

Common signs of infection include respiratory symptoms, fever, cough, shortness of breath and breathing difficulties. In more severe cases, infection can cause pneumonia, severe acute respiratory syndrome, kidney failure and even death.⁵

Standard recommendations to prevent infection spread include regular hand washing, covering mouth and nose when coughing and sneezing, thoroughly cooking meat and eggs. Avoid close contact with anyone showing symptoms of respiratory illness such as coughing and sneezing.⁵

【PRINCIPLE】

The 2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma) is a qualitative membrane-based immunoassay for the detection of IgG and IgM antibodies to 2019-nCoV in whole blood, serum or plasma specimens. This test consists of two components, an IgG component and an IgM component. In the IgG component, anti-human IgG is coated in IgG test line region. During testing, the specimen reacts with 2019-nCoV antigen-coated particles in the test cassette. The mixture then migrates upward on the membrane chromatographically by capillary action and reacts with the anti-human IgG in IgG test line region. If the specimen contains IgG antibodies to 2019-nCoV a colored line will appear in IgG test line region as a result of this. Similarly, anti-human IgM is coated in IgM test line region and if specimen contains IgM antibodies to 2019-nCoV, the conjugate-specimen complex reacts with anti-human IgM and a coloured line appears in IgM test line region as a result.

Therefore, if the specimen contains 2019-nCoV IgG antibodies, a colored line will appear in IgG test line region. If the specimen does not contain 2019-nCoV antibodies, no colored line will appear in either of the test line regions, indicating a negative result. To serve as a procedural control, a coloured line will always appear in the control line region, indicating that the proper volume of specimen has been added and membrane wicking has occurred.

【REAGENTS】

The test contains anti-human IgM and anti-human IgG as the capture reagent, 2019-nCoV antigen as the detection reagent. A goat anti-mouse IgG is employed in the control line system.

【PRECAUTIONS】

- For professional *in vitro* diagnostic use only. Do not use after expiration date.
- Do not eat, drink or smoke in the area where the specimens or kits are handled.
- Do not use the test if the pouch is damaged.
- Handle all specimens as if they contain infectious agents. Observe established precautions against microbiological hazards throughout all procedures and follow the standard procedures for proper disposal of specimens.
- Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.
- Please ensure that an appropriate amount of sample is used for testing. Too much or too little sample size may lead to erroneous results.
- The used test should be discarded according to local regulations.
- Humidity and temperature can adversely affect results.

【STORAGE AND STABILITY】

Store as packaged in the sealed pouch at room temperature or refrigerated (2-30°C). The test is stable through the expiration date printed on the sealed pouch. The test must remain in the sealed pouch until use. **DO NOT FREEZE.** Do not use beyond the expiration date.

【SPECIMEN COLLECTION AND PREPARATION】

- The 2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma) can be performed using whole blood (from venipuncture or fingerstick), serum or plasma.
- To collect **Fingerstick Whole Blood Specimens:**
 - Wash the patient's hand with soap and warm water or clean with an alcohol swab. Allow to dry.
 - Massage the hand without touching the puncture site by rubbing down the hand towards the fingertip of the middle or ring finger.
 - Puncture the skin with a sterile lancet. Wipe away the first sign of blood.
 - Gently rub the hand from wrist to palm to finger to form a rounded drop of blood over the puncture site.
 - Add the Fingerstick Whole Blood specimen to the test by using a **pipette**
 - Touch the end of the pipette to the blood until filled to approximately 20µL.[above the demarcated line.] Avoid air bubbles.
 - TO COLLECT VENOUS BLOOD SPECIMENS**
 - Collect venous blood into an EDTA tube**
- Separate serum or plasma from blood as soon as possible to avoid hemolysis. Use only clear non-hemolyzed specimens.
- Testing should be performed immediately after the specimens have been collected. Do not

leave the specimens at room temperature for prolonged periods. Serum and plasma specimens may be stored at 2-8°C for up to 7 days, for long term storage, serum/plasma specimens should be kept below -20°C. Whole blood collected by venipuncture should be stored at 2-8°C if the test is to be run within 2 days of collection. Do not freeze whole blood specimens. Whole blood collected by fingerstick should be tested immediately.

- Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Specimens should not be frozen and thawed repeatedly.
- If specimens are to be shipped, they should be packed in compliance with local regulations covering the transportation of etiological agents.
- EDTA K2, Heparin sodium, Citrate sodium and Potassium Oxalate can be used as the anticoagulant for collecting the specimen.

【MATERIALS】

- Test cassettes
- Pipettes
- Package insert
- Buffer
- Lancet and swab
- Materials required but not provided
- Specimen collection containers
- Timer
- Centrifuge (for plasma or serum)

【DIRECTIONS FOR USE】

Allow the test, specimen, buffer and/or controls to reach room temperature (15-30°C) prior to testing.

- Remove the test cassette from the foil pouch and use it within one hour. Best results will be obtained if the test is performed immediately after opening the foil pouch.
- Place the cassette on a clean and level surface.

For **Serum or Plasma** specimen:

- To use a dropper: Hold the dropper vertically, draw the specimen to the **fill line** (approximately 10µL), and transfer the specimen to the specimen well (S), then add **2 drops of buffer** (approximately 80 µL), and start the timer.
- To use a pipette: To transfer **10 µL** of specimen to the specimen well(S), then add **2 drops of buffer** (approximately 80 µL), and start the timer

For **Venipuncture Whole Blood** specimen:


- To use a dropper: Hold the dropper vertically, draw the specimen about **1 cm above the fill line** and transfer **1 full drop** (approx. 20µL) of specimen to the sample well(S). Then add **2 drops of buffer** (approximately 80 µL) and start the timer.
- To use a pipette: To transfer **20 µL** of whole blood to the specimen well(S), then add **2 drops of buffer** (approximately 80 µL), and start the timer

For **Fingerstick Whole Blood** specimen:

- To use a dropper: Hold the dropper vertically, draw the specimen about **1 cm above the fill line** and transfer **1 full drop** (approx. 20µL) of specimen to the sample well(S). Then add **2 drops of buffer** (approximately 80 µL) and start the timer.
- To use a capillary tube: Fill the capillary tube and transfer **approximately 20µL of fingerstick whole blood specimen** to the specimen well (S) of test cassette, then add **2 drops of buffer** (approximately 80 µL) and start the timer. See illustration below.

- Wait for the coloured line(s) to appear. **Read results at 10 minutes.** Do not interpret the result after 20 minutes.

Note: It is suggested not to use the buffer, beyond 6 months after opening the vial.



【INTERPRETATION OF RESULTS】

IgG POSITIVE: Two coloured lines appear. One coloured line should always appear in the control line region (C) and another line should be in the IgG line region.

IgM POSITIVE: Two coloured lines appear. One coloured line should always appear in the control line region (C) and another line should be in the IgM line region.

IgG and IgM POSITIVE: Three coloured lines appear. One coloured line should always appear in the control line region (C) and two test lines should be in the IgG line region and IgM line region.

***NOTE:** The intensity of the colour in the test line regions may vary depending on the concentration of 2019-nCoV antibodies present in the specimen. Therefore, any shade of color in the test line region should be considered positive.

NEGATIVE: One coloured line appears in the control line region (C). No lines appear in the IgG region and IgM region.

INVALID: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test. If the problem persists, discontinue using the test kit immediately and contact your local distributor.

【QUALITY CONTROL】

Internal procedural controls are included in the test. A coloured line appearing in the control region (C) is an internal procedural control. It confirms sufficient specimen volume and correct procedural technique. Control standards are not supplied with this kit; however, it is recommended that positive and negative controls be tested as a good laboratory practice to confirm the test procedure and to verify proper test performance.

【LIMITATIONS】

- The 2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma) is for *in vitro* diagnostic use only. This test should be used for detection of IgG and IgM antibody to 2019-nCoV in whole blood, serum or plasma specimens. Neither the quantitative value nor the rate of increase in the concentration of IgG or IgM antibodies to 2019-nCoV can be determined by this qualitative test.
- The 2019-nCoV IgG/IgM Rapid Test Cassette (Whole blood/Serum/Plasma) will only indicate the presence of IgG and IgM antibodies to 2019-nCoV in the specimen and should not be used as the sole criteria for the diagnosis of 2019-nCoV infections.
- As with all diagnostic tests, all results must be considered with other clinical information available to the physician.
- If the test result is negative and clinical symptoms persist, additional follow-up testing using other clinical methods is suggested. A negative result at any time does not preclude the possibility of 2019-nCoV infection.
- The hematocrit level of the whole blood can affect the test results. Hematocrit levels need to be between 25% and 65% for accurate results.
- The test will show negative results under the following conditions: The titer of the novel coronavirus antibodies in the sample is lower than the minimum detection limit of the test, or the novel coronavirus antibody has not appeared at the time of sample collection (Asymptomatic stage).

【PERFORMANCE CHARACTERISTICS】

Sensitivity and Specificity

The 2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma) was compared with a leading commercial PCR; the results show that 2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma) has a high sensitivity and specificity.

IgG Result

2019-nCoV IgG/IgM Rapid Test	Method	PCR		Total Results
	Results	Positive	Negative	
	Positive	20	1	21
	Negative	0	49	49
Total Result		20	50	70

Relative Sensitivity: 100% (95%CI*: 86.0%-100%)

Relative Specificity: 98.0% (95%CI*: 89.4%-99.9%)

Accuracy: 98.6% (95%CI*: 92.3%-99.96%)

*Confidence Interval

IgM Result

2019-nCoV IgG/IgM Rapid Test	Method	PCR		Total Results
	Results	Positive	Negative	
	Positive	17	2	19
	Negative	3	48	51
Total Result		20	50	70

Relative Sensitivity: 85.0% (95%CI*: 62.1%-96.8%)

Relative Specificity: 96.0% (95%CI*: 86.3%-99.5%)

Accuracy: 92.9% (95%CI*: 84.1%-97.6%)

*Confidence Interval

Cross-reactivity

The 2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma) has been tested for anti-influenza A virus, anti-influenza B virus, anti-RSV, anti-Adenovirus, HBSAg, anti-Syphilis, anti-H. Pylori, anti-HIV and anti-HCV positive specimens. The results showed no cross-reactivity.

Interfering Substances

The following compounds have been tested using the 2019-nCoV IgG/IgM Rapid Test Cassette (Whole Blood/Serum/Plasma) and no interference was observed.

Triglyceride: 50 mg/dL

Hemoglobin 1000mg/dL

Total cholesterol : 6mmol/L

Ascorbic Acid: 20mg/dL

Bilirubin: 60mg/dL

【BIBLIOGRAPHY】

- World Health Organization (WHO). WHO Statement Regarding Cluster of Pneumonia Cases in Wuhan, China. Beijing: WHO; 9 Jan 2020. [Accessed 26 Jan 2020]. <https://www.who.int/china/news/detail/09-01-2020-who-statement-regarding-cluster-of-pneumonia-cases-in-wuhan-china>
- Weiss SR, Leibowitz JL. Coronavirus pathogenesis. Adv Virus Res 2011;81:85-164. PMID:22094080 DOI:10.1016/B978-0-12-385885-6.00009-2
- Su S, Wong G, Shi W, et al. Epidemiology, genetic recombination, and pathogenesis of coronaviruses. Trends Microbiol 2016;24:490-502. PMID:27012512 DOI:10.1016/j.tim.2016.03.003
- Cui J, Li F, Shi ZL. Origin and evolution of pathogenic coronaviruses. Nat Rev Microbiol 2019;17:181-192.PMID:30531947 DOI:10.1038/s41579-018-0118-9
- World Health Organization (WHO). Coronavirus. <https://www.who.int/health-topics/coronavirus>

ICT INTERNATIONAL

271 DE GOEDE HOOP ESTATE

VILLAGE LANE

NOORDHOEK

CAPE TOWN

TEL/FAX +27 21 7892979

Email:russell@ictdiagnostics.co.za

www.ictdiagnostics.co.za

Number: IPI-NCOV-25T
Effective Date: 2020-02-18